

Nomor : CBT/SPPH/26/MH/0790

Tanggal : 18 September 2026

**Kepada Yth,
Pimpinan/Direktur
Utama
Rekanan PT SUCOFINDO (Persero)**

Perihal : **informasi Tender**

Bersama ini disampaikan informasi kebutuhan untuk Perusahaan kami yang akan kami adakan sebagaimana terinci dalam uraian barang/pekerjaan sebagai berikut:

No	Deskripsi	Jumlah	Pemenuhan	Pengguna
1	Blender Stainless steel (preparation)	1	Desember 2026	SBU Laboratorium
2	Kinematic Viscosity ASTM D445	1	Desember 2026	Cabang Pontianak
3	Rotating Cylinder Electrode	1	Desember 2026	SBU Laboratorium

Untuk kelancaran proses pengadaan, diharapkan perusahaan saudara dapat memenuhi persyaratan berikut:

1. Peserta tender harus sudah memenuhi lampiran wajib pada **aplikasi e-procurement SUCOFINDO**.
2. Memiliki Nomor Induk Berusaha (NIB) sesuai dengan kategori barang yang akan di supply.
3. Peserta wajib klik **Ikut Pengadaan** sebelum batas waktu.
4. Peserta tender yang memenuhi persyaratan akan mendapatkan undangan aanwijzing.

Lampiran Spesifikasi CBT/SPPH/26/MH/0790

Nama barang ^{a)}	Blender Stainless Steel (Preparation)
Merk	WARING
Type	8011S
Persyaratan tambahan	- Type : WARING 8011 S - Series - Model : Laboratory Blender - Speed : Min 18000 rpm dan Maks 20000 rpm - Listrik Voltase 220-240 Volt - 50/60 Hz
Jumlah	1 Unit

Nama barang ^{a)}	Kinematic Viscosity ASTM D445
Merk	SI Analitic German/Seta/Koehler
Type	-
Persyaratan tambahan	Kinematic Viscosity Bath, 115-240V 50/60 Hz Specification : - Conforms to the specifications of ASTM D445, D6074, D6158; IP 71; ISO 3104; DIN 51550; FTM791-305; NF T 60-100 - Temperature Control Range : 20°C to 100°C (212°F) - Temperature Display Resolution : 0.01°C - Temperature Display Accuracy : ± 0.01°C - Bath Temperature Stability : ± 0.02°C - Bath Temperature Uniformity : ± 0.02°C - Thermometer Ports : 2 - Connectivity : USB - Drain Valve : 1/2" female NPT fitting (back of unit) - Viscometer ports : Seven round 2" (51 mm) Ports Included Accessories : - Test Port Covers (7)
Jumlah	1 Unit

Nama barang ^{a)}	Rotating Cylinder Electrode (RCE)
Merk	Metrohm
Type	Autolab PGSTAT204
Persyaratan tambahan	1. Autolab PGSTAT204 2. Autolab Rotating Cylinder Electrode (RCE) AUT.RCE 1,00 3. cylinder insert made of carbon steel 1018 4. 250 mL corrosion cell 5. Computer Set PC Computer with minimal configuration - Core i7 Processor - 16 Gb RAM - 4 Gb VGA - 500 Gb SSD - 22" Monitor 6. UPS 1 KVa
Gambar	
Jumlah	1 Unit